Spark ablation inductively coupled plasma mass spectrometry analysis of minor and trace elements in low and high alloy steels using single calibration curves

R. Maibusch,^a H.-M. Kuss,^a A. G. Coedo,^{*b} T. Dorado^b and I. Padilla^b

^aUniversity Duisburg, Lotharstrasse 1, 47057 Duisburg, Germany

^bCENIM (CSIC), Gregorio del Amo, 8.28040 Madrid, Spain

Received 1st March 1999, Accepted 2nd June 1999

A unidirectional high current pulse spark with a very fast rise-time, ensuring a rapid and complete transfer of energy to the sample, was used as the sampling system for the analysis of carbon steels and highly alloyed steels with the same operating conditions. The sparking operating conditions were optimised and a restrictive path was designed to decrease the quantity of eroded material reaching the plasma, in order to prevent deposition of material in the torch injector, and to minimise sampling cone blockage and drift effects. Spark ablation sampling efficiency and effectiveness of the restrictive path were evaluated. To compensate for differences in the amount of material ablated or for a variation in drift, ⁵⁷Fe and ⁵⁵Mn were used as internal standards. The calibration procedure was applied to the analysis of the elements Al, B, Co, Cu, Mn, Nb, P, Si, and V, present in the following certified reference materials: BCS (Bureau of Analysed Samples) SS-456 to SS-460 (residual series); CRMs (European Committee for Iron and Steel Standardisation) No 285-2 (Maraging steel), No 292-1 (niobium stabilised steel), No 295-1 (highly alloyed steel), and No 296-1 (jethete steel). When plotting intensity ratios (I_X/I_{IS}) versus concentration ratios (C_X/C_{IS}) linear calibration curves over the entire range of tested concentrations, with correlation coefficients better than 0.999, were obtained. Determination limits below 1 $\mu\text{g g}^{-1}$ were found and the precision was better than 2.8%. It has also been shown to determine carbon contents at concentration levels greater than 0.03% with RSD values below 3%. For the elements As, Sn, Ti, W and Zr, only present in one or two of the Standard Materials, the sensitivity was also evaluated. Furthermore, the possibility of obtaining reproducible transient signals from sparking periods of only a few seconds was demonstrated.

Introduction

When analysing solids, conventional inductively coupled plasma based spectrometry (ICP-AES and ICP-MS) has always been slowed down by the need to dissolve the sample. If a spark source, however, is coupled to the ICP torch, electrically conductive samples can be directly analysed. In spark ablation (SA) the sample acts as a cathode positioned above a tungsten counter electrode and the sparking source generates particles transported into the plasma (ICP), where atomisation and ionisation occur. Spark ablation, in the case of bulk analysis of metals, has received attention from different research groups for possible use in ICP (whether by emission spectroscopy or mass spectrometry). In its earliest version Human *et al.*¹ used a high voltage spark with ICP-AES and they found curved calibration graphs which related to the evaporation behaviour of some of the particles. Aziz *et al.*² maintained that the use of a high repetition source at medium voltage delivered a fine aerosol, which allowed the transportation and effective excitation in an analytical ICP. Lemarchand *et al.*³ showed that for ferrous alloys the power of detection of spark erosion as the sampling technique in ICP-AES is similar to that of direct spark emission spectrometry and also showed its feasibility for the direct analysis of high alloyed steels by determining nickel in some standard reference materials. Vujicic and Steffan⁴ applied the technique to the analysis of free cutting steels. They performed a comparison of spark emission, spark-ICP-AES and continuous nebulization-ICP-AES for the analysis of standard materials, concluding that the results provided by the different methods

were comparable. Coedo *et al.*⁵ reported the analytical performances obtained with spark OES, ICP with pneumatic nebulization and spark ablation ICP-OES for the determination of Al, Ti, Nb and V in low alloyed steels. The limits of detection (LOD) obtained with SA-ICP-OES were comparable to liquid nebulization and five times lower than those obtained from spark-OES. Webb *et al.*⁶ used a novel electronic spark source for direct solid sampling with ICP-OES, obtaining detection limits in the single digit ppm range. Gagean and Mermet⁷ compared ultraviolet laser ablation and spark ablation of metals and alloys for analysis by axially viewed ICP-AES, concluding that both systems provide good sensitivity.

When ICP-MS uses the preparation of a solution from the solid test sample, the procedure is always a source of contamination as well as of polyatomic interferences, produced by the elements present in the reagents; therefore, the interest in direct introduction techniques for solid samples in ICP-MS is increasing. Jiang and Houk⁸ applied an arc nebulization for elemental analysis of conducting solids by ICP-MS. Jakubowski *et al.*⁹ applied this technique to the analysis of the NIST SRMs Series 1162 to 1165, obtaining detection limits below 100 ng g^{-1} . Ivanovic *et al.*¹⁰ observed a direct linear relationship between electron multiplier counts and concentration for trace elements in a series of pure copper standards, without requiring either concentration normalisation or internal standardisation procedures. Investigations of Kuss *et al.*¹¹ have shown that spark ablation ICP-MS is a very good alternative to liquid analysis, especially for the elements that are critical to analyse because of the interferences caused by

the acid solutions (e.g., P and Si). Coedo *et al.*¹² used a SA-ICP-MS for the simultaneous determination of residual elements in low alloyed steels and found that the detection limits were comparable to those obtained with continuous nebulization, concluding that SA-ICP-MS is a valid sampling system for the simultaneous determination of P, Si, Mn, Al, V, B, Co, Zr and Sb in low alloy steels. A comparison of the performance of spark ablation ICP-MS and glow discharge mass spectrometry (GDMS) for direct element analysis of conductive solids was carried out by Jakubowski *et al.*,¹³ showing that spark ablation should therefore be considered as a promising and powerful option for the direct element analysis of solids. Borisov *et al.*¹⁴ applied the technique to the determination of V, Rh and Pt in automotive catalytic converters, obtaining very encouraging preliminary results.

The purpose of this paper is to develop a procedure for the determination of residual elements in carbon steels and highly alloyed steel by SA-ICP-MS, under the same operating conditions. One goal of the study was to design the transfer line between the sparking cell and the torch in order to decrease the quantity of material reaching the plasma, maintaining a good sensitivity and reproducibility. The improvement of precision and accuracy of the analytical data was achieved by the use of $^{57}\text{Fe}^+$ and $^{55}\text{Mn}^+$ as internal standards. Furthermore, tests were carried out to evaluate the possibilities of the system to also determine the carbon content of the selected samples.

Experimental

Instrumentation

An ELAN 6000 ICP-MS (Perkin-Elmer SCIEX, Thornhill, ON, Canada) with a SPARK 10 (Varian Techtron, Mulgrave, Victoria, Australia) as sampling device was used. The ICP-MS and the SPARK operating conditions are listed in Table 1. The spray chamber from the ICP-MS was removed and the eroded material was carried by an Ar flow through a Tygon

tube into the injector of the ICP torch. A valve system and a transfer path were applied to improve the line connection between the sparking cell and the ICP. Fig. 1 shows a scheme of the complete system with the valves system and the transfer path included.

Standards and calibration samples

Two standard samples, CENIM C-305 (carbon steel) and C-317 (stainless steel) were used for optimisation of the system. The Certified Reference Materials BCS (Bureau of Analysed Samples) SS-456 to 460 (residual series) and CRM (European Committee for Iron and Steel Standardisation) No. 285-2 (Maraging steel), No. 292-1 (niobium stabilised steel), No. 295-1 (highly alloyed steel) and No. 296-1 (jethete steel) were used within the calibration curves. Pure iron Standard Reference Material BAM 098/1 was used for background and for determination limits calculation. The samples in a disc-shaped form were finished to a clean uniform surface by using a surface grinder with 60-grit abrasive finer zirconium oxide belt in an abrading device.

Discussion

Sparking

In spark ablation ICP spectrometry, the sample serves as the cathode. A high melting-point tungsten anode is used to provide a high voltage, and high frequency, unidirectional spark. The material ablated from the cathode is transported by an argon flow to the ICP where the ionisation takes place. The spark device used was developed firstly for application in ICP-AES, and to be applied in ICP-MS some modifications had to be made in order to minimise the mass of eroded material reaching the plasma along an entire sampling sequence (pre-spark, read time delay and measurement), thus minimising the risk of cone blockage and drift effects.

To satisfy the above objective two main modifications were introduced: namely, a valve system and a restrictive path.

The valve system (controlled manually, or automatically by the Spark Timer) was set up in order to make possible the separation of the pre-spark gas stream from the spark gas stream. With this modification, during the pre-sparking time (St1) the eroded material flows to the waste, and only during the analyses time (St2) does the sample reach the plasma. In order to maintain the plasma conditions constant during the pre-sparking time, an additional Ar stream (with a flow rate similar to that used for sampling transport) was introduced through the nebulizer flow. Furthermore, this system offers the possibility of regulating (for periods of a few seconds) the time during which the sample reaches the plasma, thus providing a transient signal. At the end of each sparking process a flushing step was applied to clean the system; during this step the gas flow is again diverted to waste. Before analysing a new sample the anode was always cleaned with a wire brush.

Moreover, because of the high sensitivity of the technique and its low tolerance to the size of particles reaching the plasma, several restricted paths were tested. The effectiveness of the different restrictive devices in removing large eroded particles from the Ar stream was evaluated by trapping the particles during a fixed period of 5 min in a 5% (1+3) HCl-HNO₃ acid solution, at different points in the transfer line. Subsequently, after heating the solutions their Fe content was determined by ICP-AES. This determination was carried out before evaluating the optimum sparking operating conditions to be sure that during the optimisation process no excessive quantity of particles reached the plasma. For these preliminary tests the sparking conditions applied were: 500 Hz, 40 A and a pulse rate of 8 μs (recommended conditions in the Varian SPARK 10 manual for the analysis of low alloy steels by ICP-

Table 1 Operating conditions for ELAN 6000 and SPARK 10

<i>ELAN 6000—</i>	
Rf power/W	1200
Plasma Ar flow/l min ⁻¹	15
Nebulizer Ar flow/l min ⁻¹	0.70
Analog stage voltage	−2000
Pulse stage voltage	1350
Mass spectrometer running pressure/Pa	3.68×10^{-5}
Dwell time per analyte/ms	200
Dwell time for ^{57}Fe /ms	2
Dwell time for ^{55}Mn /ms	4
Sweeps per reading	5
Readings per replicate	1
Number of replicates	3
Analytes	$^{11}\text{B}^+$, $^{12}\text{C}^+$, $^{27}\text{Al}^+$, $^{31}\text{P}^+$, $^{28}\text{Si}^+$, $^{48}\text{Ti}^+$, $^{51}\text{V}^+$, $^{59}\text{Co}^+$, $^{63}\text{Cu}^+$, $^{75}\text{As}^+$, $^{90}\text{Zr}^+$, $^{93}\text{Nb}^+$, $^{120}\text{Sn}^+$, $^{184}\text{W}^+$
Internal standards	$^{55}\text{Mn}^+$, $^{57}\text{Fe}^+$
Integration time per analyte/ms	1000
Integration time for ^{57}Fe /ms	10
Integration time for ^{55}Mn /ms	20
Scanning mode	Peak hopping
Estimated sample reading/s	42
Read delay/s	15
<i>SPARK 10—</i>	
Pre-spark and spark conditions	Frequency, 500 Hz; current, 60 A; pulse, 8 μs
Spark transport-flow/l min ⁻¹	1.4 (axial, 1.0; auxiliary, 0.4)
Timing/s	Pre-spark, 50; spark, 40
Tip gap/mm	3

Fig. 2 W and Mn intensity dependence *versus* frequency and current.

removing the bigger particles (which cause flickering in the ICP) is much lower in comparison with the transfer path used in this work.

The burning spots show that a partial melting process occurs; further, a partial re-deposit of the ablated material at the edge of the burn spot could be observed. Fig. 3 (A, B and C) shows photographs of the spot surface in sample C-305, after applying different current values at a pre-fixed frequency of 500 Hz. As can be observed, the more energetic the spark, the more significant are the melting areas. It may be assumed that extended melting areas give rise to condensates of metallic vapor, formed by evaporation from the molten phase, that could play an uncontrolled role during the ablation process. Fig. 4 shows the SEM spectra from the outer, black zone and from the inner, central zone of the burning crater, obtained with the selected sparking parameters; as can be seen the oxygen signal is significantly higher in the black outer zone, which could be explained by the faster oxidation of the condensed metal at the edge of the burn.

The argon transport flow rate used was 1.41 min^{-1} and an additional carrier gas flow (the nebulizer gas flow of the ICP-MS, optimised by the ICP-MS software) of about 0.71 min^{-1} was added before the restriction system. Thus, the Ar flow (after the restriction) reaching the plasma was about 0.751 min^{-1} , and this provided efficient sample transport and plasma stability. By the above monitored optimised sparking conditions, carbon steel samples and highly alloyed steel samples can be analysed with good sensitivity and stability. Plots of continuous measurements for $^{12}\text{C}^+$, $^{51}\text{V}^+$, $^{55}\text{Mn}^+$, $^{57}\text{Fe}^+$ and $^{84}\text{W}^+$ are shown in Fig. 5 for samples C-305 (carbon steel, 98% Fe) and C-317 (high alloyed steel, 70% Fe). As the isotopes $^{55}\text{Mn}^+$, $^{57}\text{Fe}^+$ were measured with reduced integration time (Table 1) sensitivity cannot be compared using the plotted intensity values (counts); this comparison can only

Fig. 3 Photographs of the spot surface in sample C-305. Frequency 500 Hz and different current values. (A) 20 A; (B) 60 A; (C) 80 A.

be made when counts s^{-1} are used. For a majority of elements the signal intensities reached a steady-state after approximately 20 s; nevertheless the best stability for all the tested elements (RSD values $<2.8\%$) was obtained by using a pre-sparking time of 50 s. After the pre-sparking period, and for quantitative measurements, a time reading delay of 15 s was applied in order to allow the sample to reach the torch and the plasma to be stabilised (this reading delay is the time elapsed between the pre-spark period and the beginning of the data acquisition). According to the stabilised operating conditions (Table 1), the ICP-MS measurement period was 42 s. Thus, the duration of a SA-ICP-MS process including pre-sparking, read delay and measurement was about 107 s. As can be observed, the signal intensities still remain stable for a longer time (Fig. 5).

Mass spectra

The intensities of the background spectrum in the range m/z 10–80 were considerably lower than those of aqueous nebulized

Fig. 4 SEM spectra. A, From the middle of the spark spot; B, from the outer black zone of the spark spot.

Fig. 5 Time profile measurements for samples C-305 and C-317.

solutions containing mineral acids, especially when HCl was present. Using this direct sample introduction technique the determination of V and As is very efficient, as there is no interference from ClO^+ and ArCl^+ . Fig. 6 shows comparatively the background spectrum of those elements yielding high background values with liquid nebulization for spark ablation and for acid nebulization. Tests were carried out to study the possibility of C determination. As significant peaks from $^{12}\text{C}^+$ were observed it should be noted that it is recommended to use argon of higher purity, as C might be present as an impurity in argon used for normal ICP analysis. Nevertheless, measurements conducted with the purest Ar quality available (99.999%) indicate that the C^+ ion only partially arises from C impurity in Ar; this verifies the theory of Jakubowski *et al.*¹³ about the presence of contaminant

gases diffusing into the plasma from the atmosphere. The BEC (background equivalent concentration) value related to an Ar purity of 99.995% was about 0.1%; this value increases to 20% in the case of the purest quality argon (99.999%).

Transient signals

As was pointed out above, the valve system introduced offers the possibility of regulating (for a few seconds) the time at which the sample reaches the plasma, in order to get a transient signal. Fig. 7 shows the transient peaks obtained from sample C-305 by opening the valve for 2, 5, 10 and 15 s, respectively. It can be seen that from 2 to 5 s the peak height increases proportionally for all the elements and remains constant thereafter, but peak area continues to rise proportionally. The relative standard deviation of the Fe and Mn intensity values, obtained from four 10 s sparking spots with 10 s period delays between the spots, were 2.1 and 1.9%, respectively. The clean out time was very short and the background value was achieved in a few ms.

Internal standard

Calibration is one of the most important aspects of spark ablation ICP-MS and it is a limiting factor for quantitative analysis. If standard reference materials with physical properties and chemical matrices similar to the unknown samples are available, and the drift effects are not significant, quantitative data may be obtained from the net intensity values. When the curves are plotted using the net intensities of the analyte, the scattering of the different calibration samples may reflect either a difference in the amount of material ablated or poor reference data for the calibration samples. Degradation in accuracy and precision can be also produced by uncontrolled variations in working conditions as well as by drift effects due to the inevitable partial cone blockage. To compensate for these negative effects internal standardisation may be used. Any constituent of the sample which is present at a constant or at a known amount in each of the analysed samples can be used as an internal standard. It should also provide a reasonable intensity value, avowing that counting statistics are a limiting factor and ensuring reasonable ratio values.¹⁵ If an internal standard (IS) at a constant amount in all the samples is used (as it is in the case of using a Fe isotope as IS for analysing any analyte in carbon steel samples, Fe content > 99%) the calibration curves are plotted referring the ratio intensities of the analyte to the IS *versus* the analyte concentration. If the selected IS is not present at a constant amount in all the samples (as happens with the constituents in the high alloyed steels) its concentration must be exactly known in each sample and the calibration curves must be plotted according to the following function;¹⁶

$$I_X/I_{IS} = f C_X/C_{IS} \quad (1)$$

where I_X is the intensity of the analyte, I_{IS} the intensity of the internal standard, C_X the concentration of the analyte, and C_{IS} the concentration of the internal standard. In this work $^{57}\text{Fe}^+$ and $^{55}\text{Mn}^+$ were used as IS. Fe was selected because it is the main matrix element and consequently the most common IS used to analyse Fe alloys (especially in spark-OES). Nevertheless, when analysing high alloyed steels with Fe as the internal standard it is necessary to know the complete elemental composition of the samples in order to calculate (by difference to 100) the real Fe content. Mn was used as IS for two reasons, firstly as it is easy to analyse with any analytical technique, and secondly because it is always present in steel samples, producing intensity values allowing reasonable ratio values.

Fig. 6 Comparative acid nebulization and spark ablation background spectra (from $m/z=10$ to $m/z=80$).

Fig. 7 Transient signals from sample C-305 with different sparking times (2 s, 5 s, 10 s and 15 s).

Calibration graphs and analytical performance

For the evaluation of the proposed method, calibration graphs including all the selected CRM were performed by applying function (1). To show the effectiveness of the use of an IS, C calibration graphs obtained without IS (plotting directly intensities *versus* concentrations) and with Fe and Mn as ISs are presented in Fig. 8 (A, B, C). These calibration graphs were directly provided by the ELAN 6000 software, the horizontal lines showing the difference between the measured intensity and its correct position according to the certified contents. The calibration concentration range is bounded by the lowest and the highest content of range element in the used CRM samples. The use of Fe or Mn as IS resulted in comparable correlation coefficients, values better than 0.999 always being found over the entire range of tested concentrations. Table 3 lists the parameters of the calibration graphs (according to the function $I_X/I_{IS} = f(C_X/C_{IS})$) for all analytes certified in the CRM, using Fe as IS for Mn and Mn as IS for the rest of the analytes. The high alloy steel CRM samples 285-2, 292-1, 295-1 and 296-1 were analysed using the calibration graphs obtained with the low alloy steels series SS-456 to SS-460. The results found are listed in Table 4 and show a good agreement with the corresponding certified values. Calibration graphs could not be created for the elements W, As, Sn, Ti and Zr, because they are only certified in one or two of the CRMs. For these elements sensitivities were calculated from the ratio of the sample signal to the background signal (considering as background the signals provided by the pure iron sample

Fig. 8 Calibration curves for carbon: (A) without IS; (B) with Fe as IS; (C) with Mn as IS.

BAM-098/1) according to the function:

$$S = I_X - I_B/C_X \quad (2)$$

where I_X is the intensity of the analyte in the sample, I_B the intensity of the analyte in the pure iron, and C the concentration in $\mu\text{g g}^{-1}$ of the analyte in the sample. Table 5 lists these calculated sensitivity values, as well as the RSD values calculated from the standard deviation of the analyte intensities for $n=4$. Quantification limits (LOQ), considered as the concentration providing a signal of ten times the standard deviation of the signal intensities corresponding to the pure iron sample BAM CRM 098/1, were calculated from the slopes of the calibration graphs obtained by directly plotting intensities *versus* concentrations, or from the sensitivity values in the case of elements only certified in one or two of the CRMs. It should be mentioned that despite the high sensitivity of the

Table 3 Calibration graph parameters using the function $I_X/I_{IS} = a C_X/C_{IS} + b$: a =slope; s_a =standard deviation of the slope; b =intercept; s_b =standard deviation of the intercept; N_s =number of standards used within each calibration procedure. ^{57}Fe was used as IS for Mn, and ^{55}Mn as IS for the rest of the analytes

Isotope	a	s_a	b	s_b	N_s	Correlation coefficient
$^{55}\text{Mn}^+$ (Fe)	73.5	1.0	1.7	1.4	10	0.99920
$^{11}\text{B}^+$ (Mn)	0.90	0.01	0.001	0.0005	7	0.99978
$^{12}\text{C}^+$ (Mn)	0.73	0.014	0.093	0.027	8	0.99905
$^{27}\text{Al}^+$ (Mn)	33.25	0.50	-0.32	0.15	8	0.99931
$^{28}\text{Si}^+$ (Mn)	5.58	0.061	-0.45	0.14	10	0.99946
$^{31}\text{P}^+$ (Mn)	1.09	0.0091	0.0028	0.012	10	0.99982
$^{51}\text{V}^+$ (Mn)	43.5	0.50	-0.30	0.23	8	0.99965
$^{59}\text{Co}^+$ (Mn)	25.64	0.31	-0.25	0.15	10	0.99924
$^{63}\text{Cu}^+$ (Mn)	9.11	0.13	0.10	0.050	5	0.99950
$^{93}\text{Nb}^+$ (Mn)	86.5	0.08	1.5	0.27	7	0.99933

Table 4 Certified and found values (%m/m) for CRM high alloy steels

Element sample	CRM 285-2	CRM 292-1	CRM 295-1	CRM 296-1
Al cert.	0.1067 ± 0.0038	—	0.0203 ± 0.001	0.0275 ± 0.0019
Al found	0.101 ± 0.003	—	0.0190 ± 0.0005	0.0275 ± 0.0007
B cert.	0.0009 ± 0.0002	—	0.0018 ± 0.0001	—
B found	0.00088 ± 0.00002	—	0.00183 ± 0.00003	—
C cert.	0.0018 ^a	0.0367 ^a	0.0018 ^a	0.1166 ± 0.0015
C found	—	—	—	0.1170 ± 0.0025
Co cert.	7.7638 ^b	0.0255 ± 0.0023	0.0450 ± 0.002	0.0218 ± 0.0012
Co found	—	0.0263 ± 0.0005	0.0431 ± 0.0008	0.0199 ± 0.0004
P cert.	0.0053 ^a	0.0175 ± 0.0014	0.0167 ± 0.0011	0.0178 ± 0.0007
P found	—	0.0167 ± 0.0001	0.0168 ± 0.0003	0.0178 ± 0.0002
Si cert.	0.0117 ± 0.0017	0.402 ± 0.0104	0.418 ± 0.0175	0.242 ± 0.005
Si found	0.0135 ± 0.0003	0.390 ± 0.007	0.433 ± 0.009	0.230 ± 0.004
V cert.	—	—	0.0456 ± 0.0033	0.363 ± 0.012
V found	—	—	0.0443 ± 0.0005	0.362 ± 0.003

^aValue lower than the lowest standard in the calibration curve (from SS-456 to SS-460). ^bValue higher than the highest standard in the calibration curve (from SS-456 to SS-460).

Table 5 Sensitivity, expressed as $S = I_X - I_B/C_X$, for the elements Ti, As, Zr, Sn, and W

Isotope	Sensitivity/counts μg^{-1} g	RSD ($n=4$)
$^{48}\text{Ti}^+$	12954	2.02
$^{75}\text{As}^+$	530	0.57
$^{90}\text{Zr}^+$	13440	1.08
$^{120}\text{Sn}^+$	8778	0.30
$^{184}\text{W}^+$	7800	1.59

ICP-MS, no W contamination from the W electrode could be observed, this being in good agreement with the results obtained by Slickers¹⁷ using spark-AES. Table 6 shows the LOQ and the RSD values obtained for the samples with the lowest certified content values of the evaluated analyte, calcu-

lated from four repetitive sparkings at different locations on the sample.

Conclusions

The application of a unidirectional high current pulse spark has resulted in an effective sampling device for the ICP-MS bulk analysis of low and high alloy steels. The design of an appropriate restrictive path allowed a suitable degree of control of the uptake rate of aerosol reaching the plasma, providing at the same time good sensitivity and low drift and blockage effects. The use of the internal standard procedure for measurements, relating ratio intensities to ratio concentrations, permits the use of the same calibration curves for the analysis of different alloy grade steel samples. The linearity of response of the ratios makes it possible to obtain the calibration graph from only two calibration samples, bracketing the analyte concentration at both low and high ends. A precision better than 3% (calculated at the lowest content tested for each analyte) has been obtained and limits of determination (on the basis of ten times the standard deviation of a pure iron sample) at the sub- $\mu\text{g g}^{-1}$ level can be calculated for the majority of the elements. The possibility of producing reproducible transient signals from discrete sampling processes has been shown. This process increases the sample throughput and the long-term stability. Solid sampling of limited sample quantities can be effective in minimising the risk of drift and instability associated with sample deposition on the sampling cone aperture. Further work relating to the use of transient signals for quantitative determinations in spark ablation ICP-MS will be carried out.

This work was supported financially by the European

Table 6 LOQ ($10 \times s_b$) and precision (RSD $n=4$, for the lowest certified value)

Element	LOQ/ $\mu\text{g g}^{-1}$	RSD (sample)
Al	0.55	2.8 (SS-457/1)
As	0.018	0.2 (CRM 295-1)
B	0.60	1.9 (SS-456/1)
C	250	1.9 (CRM 292-1)
Co	0.100	2.1 (SS-460/1)
Cu	0.20	0.75 (CRM 285-2)
Nb	0.020	2.3 (C-317)
P	2.0	0.56 (CRM 285-2)
Si	1.0	2.14 (CRM 285-2)
Sn	0.075	0.2 (CRM 295-1)
Ti	0.015	2.6 (CRM 285-2)
V	0.015	0.1 (SS-456/1)
W	0.055	1.6 (C-317)
Zr	0.010	1.1 (CRM 285-2)

References

- 1 H. G. C. Human, R. H. Scott, A. R. Oakes and C. D. West, *Analyst*, 1976, **101**, 265.
- 2 A. Aziz, J. A. C. Broekaert, K. Laqua and F. Leis, *Spectrochim Acta, Part B*, 1984, **39**, 1091.
- 3 A. Lemarchand, G. Labarraque, P. Masson and J. A. C. Broekaert, *J. Anal. At. Spectrom.*, 1987, **2**, 481.
- 4 G. Vujicic and I. Steffan *Microchim. Acta*, 1990, **109**, 315.
- 5 A. G. Coedo, T. Dorado, J. L. Jimenez and I. Gutierrez, *J. Anal. At. Spectrom.*, 1992, **7**, 11.
- 6 C. Webb, C. S. Cooper III, A. T. Zander, J. T. Arnol, E. S. Lile and S. E. Anderson, *J. Anal. At. Spectrom.*, 1994, **9**, 263.
- 7 M. Gagean and J. M. Mermet, *J. Anal. At. Spectrom.*, 1997, **12**, 189.
- 8 S. J. Jiang and R. S. Houk, *Anal. Chem.*, 1986, **58**, 1739.
- 9 N. Jakubowski, I. Feldmann, B. Sack and D. Stuewer, *J. Anal. At. Spectrom.*, 1992, **7**, 121.
- 10 K. A. Ivanovic, D. M. Coleman, F. W. Kunz and D. Schuetzle, *Appl. Spectrosc.*, 1992, **46**, 894.
- 11 H.-M. Kuss, M. Müller, J. Petin, J. L. Jiménez Seco and K. Ohls, in *Studium der Möglichkeiten und Probleme der ICP-MS-Analysentechnik in der Eisen- und Stahlindustrie, Bericht EUR 15994 DE*, EGKS-EG-EAG, Brussels and Luxembourg, 1995, p. 78.
- 12 A. G. Coedo, T. Dorado and B. Fernandez, *J. Anal. At. Spectrom.*, 1995, **10**, 859.
- 13 N. Jakubowski, I. Feldmann and D. Stuewer, *Spectrochim. Acta, Part B*, **50**, 1995, 639.
- 14 O. V. Borisov, D. M. Coleman and R. O. Carter III, *J. Anal. At. Spectrom.*, 1997, **12**, 231.
- 15 K. E. Jarvis, A. L. Gray and R. S. Houk, *Handbook of Inductively Coupled Plasma Mass Spectrometry*, Blackie, London, 1992, p. 299.
- 16 *Techniques de l'Ingenieur, P3 Analyse chimique caracterization*, ISTRAL, 1993, p. 2756.7.
- 17 K. Slickers, in *Die automatische Atom-Emmissions-Spektral-analyse*, ed. K. A. Slickers, Giessen, Germany, 2nd edn. 1992, p. 330.

Paper 9/01650C